SCANNING ELECTRON MICROSCOPE: HITACHI S-4800

IMPORTANT THINGS TO REMEMBER

- 1. Ensure EDX and annular backscatter detectors are both fully retracted before moving the stage.
- 2. Do not touch sample or sample holder with your hands. Use gloves and proper glove etiquette.
- 3. Always move stage to the home position before attempting to exchange samples.
- 4. Always degauss after changing accelerating voltage or current or working distance.
- 5. Always set focus mode. UHR for WD < 4.7 mm, HR for WD > 4.7mm.
- 6. Ensure the top of the sample is just barely under the height gauge, to avoid colliding sample with pole piece.
- 7. When seating sample base plate on SEM stage, sample placement rod will be flush with the front of the sample exchange chamber.